

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01604	Calib. Date: 07/29/2019
Certification File No.:	01605	Calib. No.: 00046
Linearity File No.:	01606	Cert. Date: 07/29/2019
Solution File No.:	01629	Cert. No.: 00033
Sequential File No.:	01629	Lin. Date: 07/29/2019
		Lin. No.: 00032
		Soln. Date: 01/04/2020
		Soln. No.: 00260
		File Date: 01/04/2020

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0364
Control Solution %:	0.100%		Expires: 09/26/2020
Solution Control Lot:	18320		Bottle No.: 0434

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:07S		
Control 1 EC	0.097%	04:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:08S		
Control 2 EC	0.096%	04:09S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:10S		
Control 3 EC	0.095%	04:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: _____



Badge No.: 68

Date: 01/04/2020