

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01630	Calib. Date: 01/08/2020
Certification File No.:	01631	Calib. No.: 00047
Linearity File No.:	01632	Cert. Date: 01/08/2020
Solution File No.:	01637	Cert. No.: 00034
Sequential File No.:	01637	Lin. Date: 01/08/2020
		Lin. No.: 00033
		Soln. Date: 02/02/2020
		Soln. No.: 00262
		File Date: 02/02/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	18320	Expires: 09/26/2020
		Bottle No.: 0595

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:57S		
Control 1 EC	0.100%	04:57S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	04:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:58S		
Control 2 EC	0.095%	04:59S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:00S		
Control 3 EC	0.095%	05:01S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: _____



Badge No.: 68

Date: 02/02/2020