

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01649	Calib. Date:	06/10/2020
Certification File No.:	01650	Cert. Date:	06/10/2020
Linearity File No.:	01651	Lin. Date:	06/10/2020
Solution File No.:	01671	Soln. Date:	10/16/2020
Sequential File No.:	01671	File Date:	10/16/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0364
Solution Control Lot:	20320	Expires:	07/06/2022
		Bottle No.:	0998

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:25D		
Control 1 EC	0.098%	11:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:26D		
Control 2 EC	0.099%	11:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28D		
Control 3 EC	0.099%	11:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature: _____



PH. #585

Badge No.: 585

Date: 10/16/2020