

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01677 **Calib. Date:** 11/19/2020 **Calib. No.:** 00050
Certification File No.: 01678 **Cert. Date:** 11/19/2020 **Cert. No.:** 00037
Linearity File No.: 01679 **Lin. Date:** 11/19/2020 **Lin. No.:** 00036
Solution File No.: 01682 **Soln. Date:** 12/10/2020 **Soln. No.:** 00277
Sequential File No.: 01682 **File Date:** 12/10/2020

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWJ S3-0363
Control Solution %: 0.100% **Expires:** 07/06/2022
Solution Control Lot: 20320 **Bottle No.:** 0994

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:09S		
Control 1 EC	0.098%	11:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:11S		
Control 2 EC	0.096%	11:11S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:12S		
Control 3 EC	0.097%	11:13S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BROWN

First Name: KYLE

MI: W

Signature: _____

Badge No.: 089

Date: 12/10/2020