

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01649	Calib. Date: 06/10/2020
Certification File No.:	01650	Cert. No.: 00048
Linearity File No.:	01651	Cert. Date: 06/10/2020
Solution File No.:	01662	Lin. No.: 00035
Sequential File No.:	01662	Lin. Date: 06/10/2020
		Soln. No.: 00271
		Soln. Date: 08/10/2020
		File Date: 08/10/2020

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0364
Control Solution %:	0.100%		Expires: 09/26/2020
Solution Control Lot:	18320		Bottle No.: 0774

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:53D		
Control 1 EC	0.099%	09:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:54D		
Control 2 EC	0.099%	09:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56D		
Control 3 EC	0.099%	09:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature: PH 

97

Badge No.: 597

Date: 08/10/2020