

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01649 Calib. Date: 06/10/2020 Calib. No.: 00048
Certification File No.: 01650 Cert. Date: 06/10/2020 Cert. No.: 00035
Linearity File No.: 01651 Lin. Date: 06/10/2020 Lin. No.: 00034
Solution File No.: 01666 Soln. Date: 09/17/2020 Soln. No.: 00273
Sequential File No.: 01666 File Date: 09/17/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364
Control Solution %: 0.100% Expires: 07/06/2022
Solution Control Lot: 20320 Bottle No.: 1198

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:31D		
Control 1 EC	0.099%	09:32D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:33D		
Control 2 EC	0.098%	09:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:34D		
Control 3 EC	0.099%	09:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCQUADE

First Name: EDWARD

MI: J

Signature: _____



Badge No.: 597

Date: 09/17/2020