

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01118	Calib. Date: 07/09/2014
Certification File No.:	01119	Calib. No.: 00032
Linearity File No.:	01120	Cert. Date: 07/09/2014
Solution File No.:	01158	Cert. No.: 00020
Sequential File No.:	01158	Lin. Date: 07/09/2014
		Lin. No.: 00020
		Soln. Date: 11/27/2014
		Soln. No.: 00165
		File Date: 11/27/2014

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWJ S3-0363
Control Solution %: 0.100%		Expires: 08/22/2016
Solution Control Lot: 14H132		Bottle No.: 1263

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:04S		
Control 1 EC	0.099%	03:05S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:06S		
Control 2 EC	0.098%	03:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:07S		
Control 3 EC	0.099%	03:08S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCMAHON

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 80

Date: 11/27/2014