

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00953	Calib. Date: 09/18/2012
Certification File No.:	00954	Calib. No.: 00028
Linearity File No.:	00955	Cert. Date: 09/18/2012
Solution File No.:	00973	Cert. No.: 00016
Sequential File No.:	00973	Lin. Date: 09/18/2012
		Lin. No.: 00016
		Soln. Date: 12/13/2012
		Soln. No.: 00129
		File Date: 12/13/2012

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0364
Control Solution %:	0.100%		Expires: 03/02/2014
Solution Control Lot:	12C100		Bottle No.: 0777

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:43S		
Control 1 EC	0.102%	13:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:44S		
Control 2 EC	0.100%	13:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46S		
Control 3 EC	0.100%	13:46S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____

Sgt. T. Tarr

Badge No.: 427

Date: 12/13/2012