

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01169 Calib. Date: 01/08/2015 Calib. No.: 00033
Certification File No.: 01170 Cert. Date: 01/08/2015 Cert. No.: 00021
Linearity File No.: 01171 Lin. Date: 01/08/2015 Lin. No.: 00021
Solution File No.: 01190 Soln. Date: 02/27/2015 Soln. No.: 00169
Sequential File No.: 01190 File Date: 02/27/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0363
Control Solution %: 0.100% Expires: 08/22/2016
Solution Control Lot: 14H132 Bottle No.: 1275

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 10:15S | | |
| Control 1 EC | 0.102% | 10:16S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.102% | 10:16S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:17S | | |
| Control 2 EC | 0.099% | 10:18S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.102% | 10:18S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:19S | | |
| Control 3 EC | 0.100% | 10:19S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 10:19S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:20S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: MARYLOUISE

MI: -

Signature: 

Badge No.: 431

Date: 02/27/2015