

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE	Calib. No.:	00031
Calibration File No.:	01074	Calib. Date:	02/06/2014
Certification File No.:	01075	Cert. Date:	02/06/2014
Linearity File No.:	01076	Lin. Date:	02/06/2014
Solution File No.:	01104	Soln. Date:	05/30/2014
Sequential File No.:	01104	File Date:	05/30/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0364
Solution Control Lot:	13G115	Expires:	07/18/2015
		Bottle No.:	0009

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:56D		
Control 1 EC	0.101%	11:56D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:57D		
Control 2 EC	0.099%	11:58D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:58D		
Control 3 EC	0.100%	11:59D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: BROWN

First Name: KYKE

MI: W

Signature: \_\_\_\_\_



Badge No.: 089

Date: 05/30/2014