

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0382  
Location: WEST WINDSOR POLICE  
Calibration File No.: 01118 Calib. Date: 07/09/2014 Calib. No.: 00032  
Certification File No.: 01119 Cert. Date: 07/09/2014 Cert. No.: 00020  
Linearity File No.: 01120 Lin. Date: 07/09/2014 Lin. No.: 00020  
Solution File No.: 01141 Soln. Date: 09/27/2014 Soln. No.: 00162  
Sequential File No.: 01141 File Date: 09/27/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0363  
Control Solution %: 0.100% Expires: 07/18/2015  
Solution Control Lot: 13G115 Bottle No.: 0993

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:37D		
Control 1 EC	0.100%	16:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:38D		
Control 2 EC	0.099%	16:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:39D		
Control 3 EC	0.099%	16:40D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


## Changed By:

Last Name: MCMAHON

First Name: MICHAEL

MI: J

Signature: \_\_\_\_\_



Badge No.: 80

Date: 09/27/2014