

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0382  
Location: WEST WINDSOR POLICE  
Calibration File No.: 01316 Calib. Date: 03/28/2016 Calib. No.: 00037  
Certification File No.: 01317 Cert. Date: 03/28/2016 Cert. No.: 00024  
Linearity File No.: 01318 Lin. Date: 03/28/2016 Lin. No.: 00024  
Solution File No.: 01320 Soln. Date: 04/08/2016 Soln. No.: 00195  
Sequential File No.: 01320 File Date: 04/08/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0363  
Control Solution %: 0.100% Expires: 04/13/2017  
Solution Control Lot: 15D138 Bottle No.: 0160

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:44D		
Control 1 EC	0.099%	13:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:45D		
Control 2 EC	0.099%	13:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46D		
Control 3 EC	0.100%	13:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: BROWN

First Name: KYLE

MI: W

Signature: \_\_\_\_\_

*Kyle Brown # 89*

Badge No.: 89

Date: 04/08/2016