

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01316	Calib. Date: 03/28/2016
Certification File No.:	01317	Calib. No.: 00037
Linearity File No.:	01318	Cert. Date: 03/28/2016
Solution File No.:	01329	Cert. No.: 00024
Sequential File No.:	01329	Lin. Date: 03/28/2016
		Lin. No.: 00024
		Soln. Date: 05/03/2016
		Soln. No.: 00196
		File Date: 05/03/2016

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0363
Control Solution %:	0.100%		Expires: 03/22/2018
Solution Control Lot:	201603L		Bottle No.: 0868

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:23D		
Control 1 EC	0.098%	12:23D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:24D		
Control 2 EC	0.099%	12:24D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:25D		
Control 3 EC	0.098%	12:26D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:26D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BROWN

First Name: KYLE

MI: W

Signature: 

Badge No.: 089

Date: 05/03/2016