

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
 Location: WEST WINDSOR POLICE
 Calibration File No.: 01316 Calib. Date: 03/28/2016 Calib. No.: 00037
 Certification File No.: 01317 Cert. Date: 03/28/2016 Cert. No.: 00024
 Linearity File No.: 01318 Lin. Date: 03/28/2016 Lin. No.: 00024
 Solution File No.: 01357 Soln. Date: 07/19/2016 Soln. No.: 00199
 Sequential File No.: 01357 File Date: 07/19/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0363
 Control Solution %: 0.100% Expires: 03/22/2018
 Solution Control Lot: 201603L Bottle No.: 0866

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:17D		
Control 1 EC	0.099%	14:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:18D		
Control 2 EC	0.099%	14:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:19D		
Control 3 EC	0.099%	14:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MOHR

First Name: DANNY

MI:

Signature: _____



Badge No.: 471

Date: 07/19/2016