

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01445	Calib. Date: 07/26/2017
Certification File No.:	01446	Calib. No.: 00041
Linearity File No.:	01447	Cert. Date: 07/26/2017
Solution File No.:	01470	Cert. No.: 00028
Sequential File No.:	01470	Lin. Date: 07/26/2017
		Lin. No.: 00027
		Soln. Date: 11/14/2017
		Soln. No.: 00224
		File Date: 11/14/2017

Calibrating Unit:	4	Model No.: CU-34	Serial No.: DDWJ S3-0364
Control Solution %:	0.100%		Expires: 09/14/2019
Solution Control Lot:	17290		Bottle No.: 0669

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:17S		
Control 1 EC	0.102%	11:18S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	11:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:18S		
Control 2 EC	0.101%	11:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	11:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:20S		
Control 3 EC	0.101%	11:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: _____



Badge No.: 68

Date: 11/14/2017