

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01222	Calib. Date: 06/10/2015
Certification File No.:	01223	Calib. No.: 00034
Linearity File No.:	01224	Cert. Date: 06/10/2015
Solution File No.:	01273	Cert. No.: 00022
Sequential File No.:	01273	Lin. Date: 06/10/2015
		Lin. No.: 00022
		Soln. Date: 11/15/2015
		Soln. No.: 00187
		File Date: 11/15/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	15D138	Expires: 04/13/2017
		Bottle No.: 0158

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:51S		
Control 1 EC	0.101%	03:52S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:53S		
Control 2 EC	0.100%	03:53S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	03:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:54S		
Control 3 EC	0.101%	03:55S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	03:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCMAHON

First Name: MICHAEL

MI: J

Signature: _____



Badge No.: 80

Date: 11/15/2015