

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01445 Calib. Date: 07/26/2017 Calib. No.: 00041
Certification File No.: 01446 Cert. Date: 07/26/2017 Cert. No.: 00028
Linearity File No.: 01447 Lin. Date: 07/26/2017 Lin. No.: 00027
Solution File No.: 01449 Soln. Date: 08/24/2017 Soln. No.: 00218
Sequential File No.: 01449 File Date: 08/24/2017

Calibrating Unit: 4 Model No.: CU-34 Serial No.: DDWJ S3-0364
Control Solution %: 0.100% Expires: 10/28/2018
Solution Control Lot: 16330 Bottle No.: 0593

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:55D		
Control 1 EC	0.102%	11:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:56D		
Control 2 EC	0.101%	11:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	11:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:57D		
Control 3 EC	0.101%	11:58D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature: PH.  # 585

Badge No.: 585

Date: 08/24/2017