

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0382  
Location: WEST WINDSOR POLICE  
Calibration File No.: 01604 Calib. Date: 07/29/2019 Calib. No.: 00046  
Certification File No.: 01605 Cert. Date: 07/29/2019 Cert. No.: 00033  
Linearity File No.: 01606 Lin. Date: 07/29/2019 Lin. No.: 00032  
Solution File No.: 01613 Soln. Date: 09/26/2019 Soln. No.: 00255  
Sequential File No.: 01613 File Date: 09/26/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364  
Control Solution %: 0.100% Expires: 09/26/2020  
Solution Control Lot: 18320 Bottle No.: 0785

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:42D		
Control 1 EC	0.100%	14:42D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:43D		
Control 2 EC	0.100%	14:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:44D		
Control 3 EC	0.100%	14:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature:  # 585

Badge No.: 585

Date: 09/26/2019