

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	01471	Calib. Date:	12/04/2017
Certification File No.:	01472	Cert. No.:	00042
Linearity File No.:	01473	Cert. Date:	12/04/2017
Solution File No.:	01486	Lin. No.:	00028
Sequential File No.:	01486	Soln. Date:	01/28/2018
		Soln. No.:	00227
		File Date:	01/28/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0363
Solution Control Lot:	17290	Expires:	09/14/2019
		Bottle No.:	0665

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:53S		
Control 1 EC	0.097%	10:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:54S		
Control 2 EC	0.097%	10:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	10:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56S		
Control 3 EC	0.097%	10:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: _____

Badge No.: 68

Date: 01/28/2018