

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	01513	Calib. Date: 04/27/2018
Certification File No.:	01514	Calib. No.: 00043
Linearity File No.:	01515	Cert. Date: 04/27/2018
Solution File No.:	01547	Cert. No.: 00030
Sequential File No.:	01547	Lin. Date: 04/27/2018
		Lin. No.: 00029
		Soln. Date: 09/21/2018
		Soln. No.: 00236
		File Date: 09/21/2018

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	17290	Expires: 09/14/2019
		Bottle No.: 0663

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:45D		
Control 1 EC	0.098%	18:46D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:46D		
Control 2 EC	0.098%	18:47D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	18:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:48D		
Control 3 EC	0.098%	18:48D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LATORRE

First Name: FRANCESCO

MI: -

Signature: _____

Badge No.: 68

Date: 09/21/2018