

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 01548 Calib. Date: 09/26/2018 Calib. No.: 00044
Certification File No.: 01549 Cert. Date: 09/26/2018 Cert. No.: 00031
Linearity File No.: 01550 Lin. Date: 09/26/2018 Lin. No.: 00030
Solution File No.: 01555 Soln. Date: 10/28/2018 Soln. No.: 00238
Sequential File No.: 01555 File Date: 10/28/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364
Control Solution %: 0.100% Expires: 09/14/2019
Solution Control Lot: 17290 Bottle No.: 0200

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:28S		
Control 1 EC	0.097%	09:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:30S		
Control 2 EC	0.097%	09:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:31S		
Control 3 EC	0.097%	09:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LAROCCA

First Name: NICHOLAS

MI: M

Signature: 

Badge No.: 85

Date: 10/28/2018